Setups

DSD-201 Endoscope reprocessor Setup	
1	Load Disinfectant from Basin
2	Set Date
3	Set Time
4	Display Software Version
5	Input Programme
6	Water Line Disinfect
7	Disinfectant Warning Inhibit
8	Display Log Information
9	Disable Logging
10	Clear Log
11	Clear Disinfectant Cycle Count
13	Display Temperatures
14	Set Heater On Time
15	Set Heater Off Time
16	Display Disinfectant Cycle Count
17	Display Time Remaining
18	Display State Time
21	Print Entire Log
25	Print Last Run
28	Set Delayed Start Time
29	Set Delayed Start Enable
33	Set Automatic Printing Enable
88	Enter Diagnostics Menu: input code 135